

ezSnare

POLYPECTOMY SNARE



STERILE EO



R_x only

FEDERAL (USA) LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN

EN - INSTRUCTIONS FOR USE



Endo-Therapeutics, Inc.
15251 Roosevelt Blvd. Suite 204
Clearwater, FL 33760 USA

EndoTherapeutics.com
Toll Free: **+1.844.400.ENDO** (3636)

Think Within.



EN – English

<small>REF</small>	ET1800 ezSnare Polypectomy Snare, Oval, Small
<small>REF</small>	ET1801 ezSnare Polypectomy Snare, Oval, Medium
<small>REF</small>	ET1802 ezSnare Polypectomy Snare, Oval, Large
<small>REF</small>	ET1803 ezSnare Polypectomy Snare, Oval, X-Large
<small>REF</small>	ET1804 ezSnare Polypectomy Snare, Oval, Small
<small>REF</small>	ET1805 ezSnare Polypectomy Snare, Oval, Medium
<small>REF</small>	ET1806 ezSnare Polypectomy Snare, Oval, Large
<small>REF</small>	ET1807 ezSnare Polypectomy Snare, Oval, X-Large
<small>REF</small>	ET1814 ezSnare Polypectomy Snare, Hexagonal, Large

Intended Use

Polypectomy snares are indicated for use in the removal, via electrocautery, of polyps during endoscopic surgery.

Contraindications

None known within the scope of the intended use.

Warnings and Precautions

This Device is Sharp! Use same care as for any other surgical device to avoid punctures or cuts to hands, gloves, etc.

Do not use if the sterile packaging is compromised, if any abnormalities are found or if the instrument is not operating properly.

Do not kink tubing when loading into endoscope.

Contact of the snare wire with the endoscope during the procedure may cause grounding, which can lead to injury of the patient and/or operator and damage to the endoscope.

Connection to equipment not complying with EN 60601-2-2 could compromise patient safety.

It is important to use an OLYMPUS® compatible plug. Other plug configurations can lead to faulty operation.

Max peak high frequency output voltage for this accessory device is 4.5 kVp-p.

Instructions

1. Peel open the pouch and remove the device. Visually inspect the device for kinks, bends, breaks or fraying and inspect the handle to ensure there are no broken parts or exposed metal. Check the hot plug and confirm that the prongs are straight and aligned. If any abnormality is found, do not use the device.
2. Inspect the active cord for kinks, bends, breaks and exposed wires. If any abnormality is found, do not use the cord.
3. Confirm the electrosurgical unit is in the OFF position. Attach an Olympus® compatible active cord to the instrument and ensure the connection is tight. The cord should snap onto the handle. If the cord is loose, call your sales representative who can provide you the proper cord.
4. Connect the other end of the cord to the electrosurgical generator and follow the manufacturer's instructions regarding the operation and proper grounding.

NOTE: IT IS IMPORTANT TO USE AN OLYMPUS® COMPATIBLE PLUG. OTHER PLUG CONFIGURATIONS CAN LEAD TO FAULTY OPERATION.

5. Fully retract the snare loop into the sheath prior to inserting the device into the working channel of the endoscope. Use short motions when inserting to avoid kinking.
6. If the endoscope is equipped with an elevator, raise the elevator fully before loading.
7. When the snare is in endoscopic view, extend the loop from the sheath and properly position the snare around the polyp.
8. Following the manufacturer's recommendations, turn on the electrosurgical generator and proceed with the polypectomy.
9. Upon completion of the procedure, turn the electrosurgical generator to the OFF position and then remove the device from the endoscope with the snare loop in the fully retracted position. Failure to fully retract the snare may cause damage to the endoscope.
10. Disconnect the active cord from the device. If the endoscope is equipped with an elevator, it must be lowered before removing the device.

Notes

Use of this device restricted to a trained healthcare professional.

In the event of a malfunction, discontinue use and save for return and contact your local representative.

Serious incidents that have occurred in relation to this medical device should be reported to the manufacturer and competent authority in the country where the incident occurred.

Disposal

Dispose of the device after use in accordance with accepted medical practice and applicable local, state and federal laws and regulations.

Storage



Store sterile instruments at room temperature.

Do not expose to disinfectants, solvents, radiation, UV light or any other substance or environment that could compromise the sterility of the unit.

Do not place any objects on the device or packaging.

Symbols

EN - Symbols

	Quantity of devices in pouch
	Quantity of devices in box

Think Within.



Endo-Therapeutics, Inc.
15251 Roosevelt Blvd, Suite 204
Clearwater, FL 33760
USA

EndoTherapeutics.com
Toll Free: **+1.844.400.ENDO** (3636)

EndoT® and Think Within.® are registered trademarks of Endo-Therapeutics, Inc.
©2018-2024 Endo-Therapeutics, Inc. All rights reserved. P/N 151204, Rev. F, 05/07/2024