

EndoTBasket

RETRIEVAL DEVICE



STERILE EO



Rx only

FEDERAL (USA) LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN

EN - INSTRUCTIONS FOR USE



Endo-Therapeutics, Inc.
15251 Roosevelt Blvd. Suite 204
Clearwater, FL 33760 USA

EndoTherapeutics.com
Toll Free: +1.844.400.ENDO (3636)

Think Within.



Intended Use

The EndoTBasket is intended for use in the endoscopic retrieval of foreign bodies.

Contraindications

None known within the scope of the intended use.

Warnings and Precautions

Do not use if package is opened or damaged.

Visually inspect before use. Do not use if any abnormality is found or if the device is not operating properly.

This device is designed for single use only. Attempts to reprocess, resterilize, and/or reuse may lead to device failure and/or transmission of disease.

Patient clinical conditions and type of foreign bodies may result in using various retrieval techniques. The technique used is at the discretion of the physician.

Do not kink tubing when loading into endoscope.

Avoid blindly passing the device past the foreign body.

Avoid inadvertently grasping tissues or organs not intended for retrieval.

Do not use to retrieve sharp objects.

Do not deploy handle quickly, in an acute position or when the endoscope is in an extremely articulated or coiled position.

Maintain moderate pressure when passing the trachea to avoid losing the foreign body.

Use caution when removing the foreign body after retrieval. Not doing so may result in damage to the device.

Instructions

1. Verify the endoscope accessory channel size is compatible with the device.
2. Peel open the pouch, remove the device from pouch and remove packaging tubing from the device handle. Visually inspect the device for kinks, loose or broken parts.
3. Form the device into a “U” shape and deploy and retract to verify proper actuation.
4. Retract the device into the outer sheath before inserting the device into the working channel of the endoscope. If resistance is met when the bending tip of the endoscope is angulated, release the bending angle slightly until the instrument passes smoothly. Do not force the device into an angulated endoscope as this could damage the endoscope.
5. Deploy, position and retract the device to retrieve the foreign body.
6. After the foreign body is collected, retract the device and endoscope while maintaining moderate pressure when passing the trachea to avoid losing the foreign body.

Notes

Use of this device restricted to a trained healthcare professional.

In the event of a malfunction, discontinue use and save for return and contact your local representative.

Serious incidents that have occurred in relation to this medical device should be reported to the manufacturer and competent authority in the country where the incident occurred.

Disposal

Dispose of the device after use in accordance with accepted medical practice and applicable local, state and federal laws and regulations.

Storage



Store sterile instruments at room temperature.

Do not expose to disinfectants, solvents, radiation, UV light or any other substance or environment that could compromise the sterility of the unit.

Do not place any objects on the device or packaging.

Symbols

EN - Symbols

	EN - Quantity of devices in pouch
	EN - Quantity of devices in box

Think Within.



Endo-Therapeutics, Inc.
15251 Roosevelt Blvd, Suite 204
Clearwater, FL 33760
USA

EndoTherapeutics.com
Toll Free: **+1.844.400.ENDO** (3636)